

Certificate

Traceability of Manufacturers to the IFCC Reference Measurement Procedure for HbA1c

This certifies that *Polymer Technology Systems Inc.* using *A1CNow+*, participates in the Monitoring Programme to demonstrate traceability. In the Monitoring Programme of 2016 the following performance was seen:

Deviation from IFCC-target	at 30 mmol HbA1c/mol Hb :	-1.4
	at 60 mmol HbA1c/mol Hb :	-0.9
	at 90 mmol HbA1c/mol Hb :	-0.4
Reproducibility, coefficient of variation		6.40%
Linearity, correlation coefficient		0.9905

* PTS Addendum: See attached statement regarding commutability of IFCC traceability samples

Date of issue: 1 December 2016

Certification expires: 31 December 2017



Dr. C.W. Weykamp
IFCC HbA1cNetwork Coordinator