

pts panels®

Multi-Chemistry Controls

For use with CardioChek® brand analyzers

INTENDED USE

For use with PTS Panels® test strips as quality control materials to check for proper performance of your test system.

SUMMARY AND EXPLANATION

PTS Panels multi-chemistry controls contain cholesterol, triglycerides, and glucose, which react with test strips to produce color or an electrical current depending on the test strips used. The result for each level of control should be in the range specified on the quality control range card included with these materials. Controls should be run to verify the test system performance, when results are questionable, to comply with your facility's quality control requirements, or as required by local accrediting and regulatory bodies.

Note: HDL cholesterol results are reported when using multi-chemistry controls (Lot MC24 and up) on CardioChek® Plus v. 1.09 analyzers and above. HDL cholesterol controls are required for all other CardioChek brand analyzers.

CHEMICAL COMPOSITION

PTS Panels multi-chemistry controls contain glucose, glycerol, cholesterol, DL-β hydroxybutyric acid, preservatives, and inert ingredients.

PRECAUTIONS

- For *in vitro* diagnostic use. Intended for professional and self-testing use.
- Do not eat or drink contents.
- Avoid contact with eyes or skin.

STORAGE AND HANDLING

Store in a cool, dry place at room temperature of 68–86°F (20–30°C). May be stored in a refrigerator at 35–46°F (2–8°C). Do not freeze. Product is ready to use and is clear. Recap after using. Use care not to contaminate dropper tip. Do not use beyond the expiration date on the vial. Opened product is stable for at least ten months at room temperature.

MATERIALS PROVIDED

- Two (2) vials of multi-chemistry controls (one each of Level 1 and Level 2)

MATERIALS NEEDED BUT NOT PROVIDED

- CardioChek brand analyzer
- PTS Panels test strips and MEMO Chip® of same lot number

PROCEDURE


IMPORTANT: Read all instructions carefully before testing.

Multi-Chemistry Controls Test Procedure

1. Insert MEMO Chip that matches the lot of strips.
2. Press one of the two buttons to turn the instrument ON.
3. Press the NEXT button until the display reads UTILITY.
4. Press the ENTER button.
5. Press the NEXT button until RUN CONTROL is displayed.
6. At the RUN CONTROL analyzer display option, press ENTER.
7. Insert the test strip into the analyzer.
8. Wait until analyzer displays APPLY SAMPLE.
9. Invert to mix control.
10. Remove cap and turn bottle upside down straight above reflectance test strip sample application window or to the tip of the electrochemical test strip.

11. Carefully squeeze vial to deliver control solution on the test strip. Use one drop of control solution for single-analyte test strips and two drops for multi-analyte test strips and one drop of control solution to the tip of the electrochemical test strip.
12. Replace control vial cap.
13. Results will be displayed on screen.

Note: If testing PTS Panels Lipid+eGLU smart bundle test strips, skip the eGLU test by pressing and holding the NEXT button when INSERT STRIP icon for the lipid panel test strip appears.



EXPECTED RESULTS

Results for the control should be in the range specified on the quality control range card. The most up-to-date control results are available on the PTS Diagnostics website: www.ptsdiagnostics.com. If test results are above or below the range specified, check to see if the test strips or controls are expired. If they are expired, test again with unexpired strips and controls. Clean the analyzer before retesting (see user guide).

For assistance with the PTS Panels multi-chemistry controls, please contact PTS Diagnostics Customer Service (M-F, 6 a.m. - 9 p.m. US EST) or your local authorized dealer.

1-877-870-5610 (Toll-free inside the USA)

+1-317-870-5610 (Direct)

+1-317-870-5608 (Fax)

E-mail: customerservice@ptsdiagnostics.com

PTS Panels, CardioChek, and MEMO Chip are trademarks of Polymer Technology Systems, Inc.

© 2017 Polymer Technology Systems, Inc.



Polymer Technology Systems, Inc.
7736 Zionsville Road
Indianapolis, IN 46268 USA
1-317-870-5610



MDSS GmbH
Schiffgraben 41
30175 Hannover, Germany

EXPLANATION OF SYMBOLS

	Use by		Manufacturer
	Batch code		Temperature limitation
	<i>In vitro</i> diagnostic medical device		Caution
	Catalog number		Authorized representative in the European Community
	Consult instructions for use		
	Control		
	This product fulfills the requirements of European Directive 98/79/EC for <i>in vitro</i> diagnostic medical devices		